

# Techniques for Mammalian Cell Tissue Culture

Tissue culture technology has found wide application in the field of neuroscience research. Monolayer cell cultures are utilized in cytogenetic, biochemical, and molecular laboratories for research as well as diagnostic studies. The isolation and culture of numerous specific types of neural cells is described in Chapter 3. This appendix addresses general concerns and procedures needed for successful culture of mammalian cells. In most cases, cells or tissues must be grown in culture for days or weeks to obtain sufficient numbers of cells for analysis. Maintenance of cells in long-term culture requires strict adherence to aseptic technique to avoid contamination and potential loss of valuable cell lines. The first section of this unit discusses basic principles of sterile technique.

An important factor influencing the growth of cells in culture is the choice of tissue culture medium. Many different recipes for tissue culture media are available, and each laboratory must determine which medium best suits its needs. Individual laboratories may elect to use commercially prepared medium or prepare their own. Commercially available medium can be obtained as a sterile and ready-to-use liquid, in a concentrated liquid form, or in a powdered form. Besides providing nutrients for growing cells, medium is generally supplemented with antibiotics, fungicides, or both to inhibit contamination. The second section of this unit discusses medium preparation.

As cells grown in monolayer reach confluency, they must be subcultured or passaged. Failure to subculture confluent cells results in reduced mitotic index and eventually in cell death. The first step in subculturing is to detach cells from the surface of the primary culture vessel by trypsinization or mechanical means. The resultant cell suspension is then subdivided, or reseeded, into fresh cultures. Secondary cultures are checked for growth and fed periodically, and may be subsequently subcultured to produce tertiary cultures, etc. The time between passaging of cells varies with the cell line and depends on the growth rate.

The Basic Protocol describes subculturing of a monolayer culture grown in petri plates or flasks. Support Protocols describe freezing of monolayer cells, thawing and recovery of cells, counting cells using a hemacytometer, and preparing cells for transport. Alternate Protocols 1 and 2 describe the passaging and freezing of cells in suspension culture.

**CAUTION:** Radioactive, biological, and chemical substances require special handling; follow institutional, local, and national regulations.

## STERILE TECHNIQUE

It is essential that sterile technique be maintained when working with cell cultures. Aseptic technique involves a number of precautions to protect both the cultured cells and the laboratory worker from infection. The laboratory worker must realize that cells handled in the lab are potentially infectious and should be handled with caution. Protective apparel such as gloves, lab coats or aprons, and eyewear should be worn when appropriate (Knutsen, 1991). Care should be taken when handling sharp objects such as needles, scissors, scalpel blades, and glass that could puncture the skin. Sterile disposable plastic supplies may be used to avoid the risk of broken or splintered glass (Rooney and Czepulkowski, 2001).

Frequently, specimens received in the laboratory are not sterile, and cultures prepared from these specimens may become contaminated with bacteria, fungus, or yeast. The presence of microorganisms can inhibit growth, kill cell cultures, or lead to inconsistencies in test results. The contaminants deplete nutrients in the medium and may produce substances that are toxic to cells. Antibiotics and antimycotics can be used to combat potential contaminants (see Table A.3B.1). The solutions can be used to wash specimens prior to culture or to rinse contaminated cultures, and can be added to medium used for tissue culture. Antibiotics (penicillin, streptomycin, kanamycin, neomycin, or gentamycin) can be used individually or in combinations (i.e., penicillin/streptomycin/neomycin and penicillin/streptomycin/gentamycin). Likewise, fungicides such as amphotericin B (Fungizone) and mycostatin (Nystatin) can be used alone or as antibiotic/antimycotic solutions (e.g., penicillin/streptomycin/Fungizone). Particular care should be taken when using Fungizone because it is typically very toxic to cell cultures, and adequate data are not available regarding potential adverse effects on the growth of human cells. Antibiotics and antimycotics are available from a number of vendors including Invitrogen, Sigma-Aldrich, and BioWhittaker.

All materials that come into direct contact with cultures must be sterile. Sterile disposable dishes, flasks, pipets, etc., can be obtained directly from manufacturers. Reusable glassware must be washed, rinsed thoroughly, then sterilized by autoclaving or by dry heat before reusing. With dry heat, glassware should be heated 1.5 to 2 hr at 160°C to ensure sterility. Materials that may be damaged by very high temperatures can be autoclaved 20 min at 120°C and 15 psi. All media, reagents, and other solutions that come into contact with the cultures must also be sterile; media may be obtained as sterile liquids from the manufacturer, autoclaved if not heat-sensitive, or filter sterilized. Supplements can be added to media prior to filtration, or they can be added aseptically after filtration. Filters with 0.20- to 0.22- $\mu$ m pore size should be used to remove small gram-negative bacteria from culture media and solutions.

Contamination can occur at any step in handling cultured cells. Care should be taken to maintain the sterility of petri plates, pipets, and flasks that are used for tissue culture. The use of disposable, sterile culture supplies is convenient and has virtually eliminated the need to flame-sterilize instruments and vessels used for tissue culture. The initial cost of purchasing these single-use supplies outweighs the expense associated with the time and effort of washing, packaging, and sterilizing nondisposables. Good sterile technique must be followed. For example, if a sterile pipet tip should come into contact with the benchtop or other nonsterile surface, the pipet tip should be discarded and a fresh one obtained. If

**Table A.3B.1** Working Concentrations of Antibiotics and Fungicides for Mammalian Cell Culture

Additive	Final concentration
Penicillin	50–100 U/ml
Streptomycin sulfate	50–100 $\mu$ g/ml
Kanamycin	100 $\mu$ g/ml
Gentamycin	50 $\mu$ g/ml
Neomycin	50 $\mu$ g/ml
Mycostatin (Nystatin)	20 $\mu$ g/ml
Amphotericin B (Fungizone)	0.25 $\mu$ g/ml
Pen/strep/neomycin <sup>a</sup>	1 $\times$

<sup>a</sup>Usually supplied as 100 $\times$  concentrated stock solution, which is diluted to 1 $\times$  concentration in the medium.

disposables are not an option, the necks of sterile containers, such as bottles and flasks, and the tips of pipets should be passed through a flame before the pipet is introduced into the container. After pipetting, the neck of the bottle or flask should again be flamed.

Certain instruments used for tissue culture (forceps, scissors, scalpels, and in some instances, glass bottles) may require autoclave sterilization prior to initial use. The autoclave relies on pressurized steam to destroy microorganisms. Instruments to be autoclaved should be thoroughly washed and dried, then packaged if necessary to ensure that sterility will be maintained after removal from the autoclave until use. Indicator tape or autoclave bags with indicator strips should be used to document that the items have been autoclaved. The indicator tape demonstrates that the item has been autoclaved, but does not ensure sterility.

Historically, if instruments such as forceps, tweezers, scissors, or scalpels were going to be re-used to handle several sequential specimens, they were rapidly sterilized between uses by dipping in 70% alcohol and flaming. Other methods of rapid decontamination are now available. For example, the Germinator 500 (Cellpoint Scientific) is a small benchtop unit [height, 6.76 in. (~17 cm) × width, 5.25 in. (~13 cm) × depth 5.25 in. (~13 cm)] used in tissue culture labs for decontamination of metal dissecting instruments. Instruments must be clean, dry, and free of debris prior to sterilization. The instruments are inserted into a stainless steel well [inside dimensions: diameter, 2 in. (~5 cm) × depth, 4 in. (~10 cm)] filled with glass beads that are maintained at a constant temperature of 500°F (~260°C). Dry heat decontamination occurs through heat transfer from the glass beads to the instruments. The inserted parts of small instruments are dry sterilized within ~15 sec, while larger instruments may take as long as 1 min. The Germinator 500 eliminates the need for alcohol and open flames for sterilization. Due to the lack of a method for routine monitoring of sterilization by glass beads, the Germinator 500 operating manual recommends that this equipment be used for research purposes only.

Although tissue culture work can be done on an open bench if aseptic methods are strictly enforced, many labs prefer to perform tissue culture work in a room or low-traffic area reserved specifically for that purpose. At the very least, biological safety cabinets (BCSs) are recommended to protect the cultures, the laboratory worker, and the environment. Although there are three general categories of BCS (Classes I, II, and III), Class II are laminar flow safety cabinets that are favored for most tissue culture applications. The inward flow of air protects the lab worker from exposure to biohazards. Cultures are protected from airborne contamination through high-efficiency particulate air (HEPA) filtration of the air flow down across the work area (vertical laminar flow). The exhaust air also passes through a HEPA filter to protect the environment from particulate or aerosol contaminants. Class II BCSs are divided into Type A and Type B, which is further subdivided into B1, B2, and B3. The laboratory should consider the types of samples being processed and the types of potential pathogenic exposure in making a selection (see <http://www.cdc.gov/od/ohs/biosfty/bmbl4/b4aa.htm>). Manufacturer recommendations should be followed regarding routine maintenance checks on air flow and filters. For day-to-day use, the cabinet should be turned on for at least 5 min prior to beginning work. All work surfaces both inside and outside of the hood should be kept clean and disinfected daily and after each use. A 10% household bleach solution, 70% alcohol, an iodophor, a quarternary ammonium compound, or commercially available liquid disinfectants can be used (<http://www.ehs.cornell.edu/bio/cabinets.htm>, last updated December, 2000).

Some safety cabinets are equipped with ultraviolet (UV) lights for decontamination of work surfaces. However, UV lamps are not required in biological safety cabinets, and their effectiveness has been questioned (<http://www.niehs.nih.gov/odhsb/biosafe/bsc/section6.htm>; <http://www.ehs.cornell.edu/bio/cabinets.htm>). UV lamps may produce

a false sense of security as they maintain a visible blue glow long after their germicidal effectiveness is lost. Effectiveness diminishes over time as the glass tube gradually loses its ability to transmit short UV wavelengths, and may also be reduced by dust on the glass tube, distance from the work surface, temperature, and air movement. Even when the UV output is adequate, the rays must directly strike a microorganism in order to kill it; bacteria or mold spores hidden below the surface of a material or outside the direct path of the rays will not be destroyed. Another rule of thumb is that anything that can be seen cannot be killed by UV. UV lamps will only destroy microorganisms such as bacteria, viruses, and mold spores; they will not destroy insects or other large organisms. Aside from their general ineffectiveness, UV lights pose a safety hazard, because exposure can cause damage to the eyes and skin (<http://www.ehs.washington.edu/rsononion/uvlight.shtm>). A more reliable approach to eliminating contamination is the use of well-practiced microbiological procedures, good aseptic techniques, and standard operational procedures for working in a BSC, including thorough decontamination with an effective disinfectant before and after use of the BSC. The current recommendation is that work surfaces be wiped down with ethanol instead of relying on UV lamps, although some labs use the lamps in addition to ethanol wipes to decontaminate work areas. If UV lamps are used as a secondary method of disinfecting the work surface, their radiation output should be tested with a UV meter during the annual certification of the BSC to ensure that the proper intensity of light is reaching the work surface. The radiation output should be at least  $40 \mu\text{W}/\text{cm}^2$  at 254 nm when measured with a UV flux meter placed in the center of the work surface. The UV lamps should be replaced when they fall below the minimum requirements for protection (<http://www.ehs.cornell.edu/bio/cabinets.htm>).

Cultures should be visually assessed on a routine basis for evidence of contamination. Indicators in the tissue culture medium change color when contamination is present: for example, medium that contains phenol red changes to yellow because of increased acidity. Cloudiness and turbidity are also observed in contaminated cultures. Once contamination is confirmed with a microscope, infected cultures are generally discarded. Keeping contaminated cultures increases the risk of contaminating other cultures. Sometimes a contaminated cell line can be salvaged by treating it with various combinations of antibiotics and antimycotics in an attempt to eradicate the infection. In this procedure, the tissue culture medium is aspirated from the affected cultures and discarded. The cultures are rinsed with fresh “wash” medium that has been prepared by supplementing the routinely used tissue culture medium with the appropriate concentration of antibiotic and/or antimycotic (Table A.3B.1). The petri dish or flask should be gently swirled so that the cell surface is bathed in the clean medium. The wash medium can be left on the cultures for 1 to 5 min. Due to the toxicity of Fungizone and amphotericin, it is recommended that cell exposure to these agents be limited to less than 2 min. Aspirate the wash medium, replace with fresh medium, and return the cultures to the incubator. Even under the best conditions, such treatment may adversely affect cell growth, and it is often unsuccessful in ridding cultures of contaminants.

Mycoplasma are small (0.2 to 0.3  $\mu\text{m}$ ) intracellular bacteria that attach to the cell membrane, inhibit cell growth, and eventually lead to cell death. Because these parasites do not have cell walls, do not grow in colonies, and do not change the pH of the medium, they are difficult to detect visually in cultures. Mycoplasma can multiply to very high concentrations ( $10^7$  to  $10^8$  organisms/ml) and adversely affect cultures by altering cell growth characteristics, inhibiting cell metabolism, disrupting nucleic acid synthesis, inducing chromosome aberrations, changing cell membrane antigenicity, and altering transfection rates and viral susceptibility (<http://www.unc.edu/depts/tcf/mycoplasma.htm>).

Mycoplasma are spread by cross-contamination from infected cultures through aerosolization during pipetting, or via the transfer of contaminated cells or contaminated

reagents used in cell culture (<http://www.unc.edu/depts/tcf/mycoplasma.htm>). There are several methods for detecting mycoplasma in cell cultures and in cell culture reagents. In the direct method, both cultures and reagents can be plated on agar in order to grow and thereby reveal the presence of mycoplasma. Disadvantages of this method are that some strains of mycoplasma cannot be cultivated on agar and that the results can take from 2 to 4 weeks to obtain (<http://www.lerner.ccf.org/services/cell/cellculture.php>). Several indirect methods are available and should be selected based on the capabilities of the laboratory and the needs of the laboratory in terms of sensitivity, specificity, and time requirements. The DNA staining method uses Hoechst 33258 to highlight the A-T rich DNA of mycoplasma, so that mycoplasma appear as bright extranuclear spots in the cytoplasm. Results are available in 24 hr, but this method is not as sensitive as the direct method and can be difficult to interpret due to background bacterial/yeast/fungal contamination, excess debris, reduced or absent live cells, and broken nuclei from dead cells. The addition of an indicator such as Vero (Sigma-Aldrich) increases the sensitivity of this method. PCR primers that selectively amplify part of the mycoplasma DNA are also available. Kits include the ELISA mycoplasma detection kit from Roche Applied Science, the Mycoplasma Detection kit from Minerva Biolabs, the Myco Alert Mycoplasma Detection Assay by Cambrex (with results available in <20 min using the Cambrex instrument), and the VenorGem Mycoplasma PCR Detection Kit from Sigma-Aldrich.

Good aseptic technique and routine testing are the most effective methods for preventing mycoplasma contamination. Most standard antibiotics are not effective in treating mycoplasma contamination, and many laboratories prefer to dispose of contaminated cultures. In the event that the cultures are valuable and backup cultures are not available, it is imperative to attempt to salvage the contaminated cells. Agents such as Mynox Mycoplasma Elimination Reagent and Mycoplasma Removal Agent (MRA) are commercially available from ICN Flow, and usually require a single treatment to decontaminate cultures (<http://www.unc.edu/depts/tcf/mycoplasma.htm>).

## CULTURE MEDIUM PREPARATION

Choice of tissue culture medium comes from experience. An individual laboratory must select the medium that best suits the type of cells being cultured. Chemically defined media are available in liquid or powdered form from a number of suppliers. Sterile, ready-to-use medium has the advantage of being convenient, although it is more costly than other forms. Powdered medium must be reconstituted with tissue-culture-grade water according to manufacturer's directions. Distilled or deionized water is not of sufficiently high quality for medium preparation; double- or triple-distilled water or commercially available tissue culture water should be used. The medium should be filter-sterilized and transferred to sterile bottles. Prepared medium can generally be stored  $\leq 1$  month in a 4°C refrigerator. Laboratories using large volumes of medium may choose to prepare their own medium from standard recipes. This may be an economical approach, but it is time-consuming and the savings may not offset the time required.

Basic media such as Eagle minimal essential medium (MEM), Dulbecco's modified Eagle medium (DMEM), Glasgow modified Eagle medium (GMEM), and RPMI 1640 and Ham F10 nutrient mixture (e.g., Invitrogen) are composed of amino acids, glucose, salts, vitamins, and other nutrients. A basic medium is supplemented by addition of L-glutamine, antibiotics (typically penicillin and streptomycin sulfate), and usually serum to formulate a "complete medium." Where serum is added, the amount is indicated as a percentage of fetal bovine serum (FBS) or other serum. Some media are also supplemented with antimycotics, nonessential amino acids, various growth factors, and/or drugs that provide selective growth conditions. Supplements should be added to medium prior to sterilization or filtration, or added aseptically just before use.

The optimum pH for most mammalian cell cultures is 7.2 to 7.4. Adjust pH of the medium as necessary after all supplements are added. Buffers such as bicarbonate and HEPES are routinely used in tissue culture medium to prevent fluctuations in pH that might adversely affect cell growth. HEPES is especially useful in solutions used for procedures that do not take place in a controlled CO<sub>2</sub> environment.

Fetal bovine serum (FBS; sometimes known as fetal calf serum, FCS) is the most frequently used serum supplement. Calf serum, horse serum, and human serum are also used; some cell lines are maintained in serum-free medium (Freshney, 1996). Complete medium is supplemented with 5% to 30% (v/v) serum, depending on the requirements of the particular cell type being cultured. Serum is obtained frozen, then is thawed, divided into smaller portions, and refrozen until needed.

There is considerable lot-to-lot variation in FBS. Most suppliers will provide a sample of a specific lot and reserve a supply of that lot while the serum is tested for its suitability. The suitability of a serum lot depends upon the use. Frequently the ability of serum to promote cell growth equivalent to a laboratory standard is used to evaluate a serum lot. Once an acceptable lot is identified, enough of that lot should be purchased to meet the culture needs of the laboratory for an extended period of time.

Although the use of heat-inactivated serum was once preferred, it is no longer standard, or it may be standard in some labs simply because it is an established and routine practice ([http://www.unc.edu/depts/tc/tcf/tech\\_tipsHI.htm](http://www.unc.edu/depts/tc/tcf/tech_tipsHI.htm)). Historically, heat inactivation was considered necessary to destroy complement protein due to its role in cell lysis. It has since been shown that the level of complement components in commercially available FCS was only 1% to 3% of adult levels, and that no significant cell lysis was detected with the use of commercially available FCS, even when undiluted (Triglia and Linscott, 1980). In addition, many researchers prewarm FBS to 37°C, which is enough to inactivate heat-labile complement components. Improvements in the filtration of serum products have also made the use of heat-inactivated serum nonessential for most cell types. In the past serum was filtered through 0.45-μm or 0.22-μm filters, raising concern that adventitious agents such as mycoplasma could persist. Today media suppliers use 0.1-μm or 0.04-μm porosity membranes to eliminate the need for heat inactivation ([http://www.unc.edu/depts/tc/tcf/tech\\_tipsHI.htm](http://www.unc.edu/depts/tc/tcf/tech_tipsHI.htm); Hyclone Labs, 1996). Not only is heat inactivation unnecessary for most cell types, it may sometimes be detrimental to cell growth. Heat inactivation not only destroys complement, it also degrades amino acids, vitamins, growth factors, and other nutrients that enhance cell growth. Furthermore, heat inactivation can increase the formation of precipitates that can be mistaken for microbial contamination (Hyclone Labs, 1996). Laboratories that continue the use of heat inactivation should conduct studies to determine whether this step is really indicated. Heat activation may be warranted when conducting immunologic studies or when culturing embryonic stem cells, insect cells, and smooth muscle cells. For cells obtained from a commercial vendor such as ATCC, serum requirements are included in the cell descriptions.

Although FBS has historically been the serum of choice, many investigators are moving towards the use of alternative types of animal serum or to serum-free media. An increasing number of companies are offering serum-free media for a wide range of cell types. Focus on Alternatives (FOA) lists over 300 types of serum-free products along with their suppliers, applications, and the cell types supported for each product (<http://www.focusonalternatives.org.uk>, compiled in July, 2005). Although the term “serum-free” implies that a medium contains no serum, the medium may not be entirely free of serum-derived products. For example, bovine serum albumin may be used as the protein component of particular types of serum-free media (Newman, 2003). While some media are designed for culturing a particular cell type, others are general-purpose media that can support a variety of cell types.

Commercially prepared media containing L-glutamine are available, but many laboratories choose to obtain medium without L-glutamine, and then add it to a final concentration of 2 mM just before use. L-Glutamine is an unstable amino acid that, upon storage, converts to a form cells cannot use. Breakdown of L-glutamine is temperature- and pH-dependent. At 4°C, 80% of the L-glutamine remains after 3 weeks, but at near incubator temperature (35°C) only half remains after 9 days (Barch et al., 1991). To prevent degradation, 100× L-glutamine should be stored frozen in aliquots until needed.

Another option to prevent degradation is the use of a stabilized dipeptide form of L-glutamine such as GlutaMAX (Invitrogen), which is available both as a stand-alone medium and as a medium supplement. GlutaMAX media contain the dipeptide L-alanyl-L-glutamine, which is split by aminopeptidases in culture to release L-glutamine and L-alanine. This mechanism provides a controlled delivery of L-glutamine to the cells in culture. The GlutaMAX dipeptide is also available as a medium supplement to be used as a direct substitute for L-glutamine in cell culture medium preparations with minimal or no adaptation. Both GlutaMAX media and supplement are reported by the manufacturer to maximize cell performance, improve cell viability and growth, extend cell culture life, and minimize toxic ammonium build-up during cell culture (<http://www.invitrogen.com>).

In addition to practicing good aseptic technique, most laboratories add antimicrobial agents to medium to further reduce the risk of contamination. A combination of penicillin and streptomycin is the most commonly used antibiotic additive; kanamycin and gentamycin are used alone. Mycostatin and amphotericin B are the most commonly used fungicides (Rooney and Czepulkowski, 2001). Table A.3B.1 lists the final concentrations for the most commonly used antibiotics and antimycotics. Combining antibiotics in tissue culture media can be tricky because some antibiotics are not compatible and one may inhibit the action of another. Furthermore, combined antibiotics may be cytotoxic at lower concentrations than the individual antibiotics. In addition, prolonged use of antibiotics may cause cell lines to develop antibiotic resistance. For this reason, some laboratories add antibiotics and/or fungicides to medium when initially establishing a culture but eliminate them from medium used in later subcultures.

Commercially available tissue culture media have been tested for sterility prior to release, and further testing within the research laboratory is generally not required. In clinical laboratories, sterility checks on tissue culture media are often performed as quality-control in compliance with requirements of the College of American Pathologists (CAP). A small aliquot from each lot of medium is incubated 48 hr at 37°C and monitored for evidence of contamination such as turbidity (infected medium will be cloudy) and color change (if phenol red is the indicator, infected medium will turn yellow). Any contaminated medium should be discarded.

## TRYPsinIZING AND SUBCULTURING CELLS FROM A MONOLAYER

A primary culture is grown to confluency in a 60-mm petri plate or 25-cm<sup>2</sup> tissue culture flask containing 5 ml tissue culture medium. Cells are dispersed by trypsin treatment and then reseeded into secondary cultures. The process of removing cells from the primary culture and transferring them to secondary cultures constitutes a passage, or subculture.

### Materials

Primary cultures of cells

HBSS (See recipe) without Ca<sup>2+</sup> and Mg<sup>2+</sup>, 37°C

0.25% (w/v) trypsin/0.2% EDTA solution (see recipe), 37°C

Complete medium (e.g., complete DMEM; see recipe) supplemented with 10% to 15% (v/v) fetal bovine serum, 37°C

## BASIC PROTOCOL

### Commonly Used Techniques

## A.3B.7

Sterile Pasteur pipets  
37°C warming tray or incubator  
Tissue culture plasticware or glassware including pipets and 25-cm<sup>2</sup> flasks or  
60-mm petri plates, sterile

**NOTE:** All incubations are performed in a humidified 37°C, 5% CO<sub>2</sub> incubator unless otherwise specified. Some media (e.g., DMEM) may require altered levels of CO<sub>2</sub> to maintain pH 7.4.

**NOTE:** All solutions and equipment coming into contact with cells must be sterile, and proper aseptic technique should be used accordingly.

1. Remove all medium from primary culture with a sterile Pasteur pipet. Wash adhering cell monolayer once or twice with a small volume of 37°C HBSS without Ca<sup>2+</sup> and Mg<sup>2+</sup> to remove any residual FBS that may inhibit the action of trypsin.

*Use a buffered salt solution that is Ca<sup>2+</sup>- and Mg<sup>2+</sup>-free to wash cells. Ca<sup>2+</sup> and Mg<sup>2+</sup> in the salt solution can cause cells to stick together.*

*If this is the first medium change, rather than discarding medium that is removed from primary culture, put it into a fresh dish or flask. The medium contains unattached cells that may attach and grow, thereby providing a backup culture.*

2. Add enough 37°C trypsin/EDTA solution to culture to cover adhering cell layer.
3. Place plate on a 37°C warming tray 1 to 2 min. Tap bottom of plate on the countertop to dislodge cells. Check culture with an inverted microscope to be sure that cells are rounded up and detached from the surface.

*If cells are not sufficiently detached, return plate to warming tray for an additional minute or two.*

4. Add 2 ml 37°C complete medium. Draw cell suspension into a Pasteur pipet and rinse cell layer two or three times to dissociate cells and to dislodge any remaining adherent cells. As soon as cells are detached, add serum or medium containing serum to inhibit further trypsin activity that might damage cells.

*If cultures are to be split 1/3 or 1/4 rather than 1/2, add sufficient medium such that 1 ml of cell suspension can be transferred into each fresh culture vessel.*

5. Add an equal volume of cell suspension to fresh plates or flasks that have been appropriately labeled.

*Alternatively, cells can be counted using a hemacytometer (Support Protocol 3) or Coulter counter and diluted to the desired density so a specific number of cells can be added to each culture vessel. A final concentration of  $\sim 5 \times 10^4$  cells/ml is appropriate for most subcultures.*

*For primary cultures and early subcultures, 60-mm petri plates or 25-cm<sup>2</sup> flasks are generally used; larger petri plates or flasks (e.g., 150-mm plates or 75-cm<sup>2</sup> flasks) may be used for later subcultures.*

*Cultures should be labeled with at least two unique identifiers, as well as date of culture and passage number.*

6. Add 4 ml fresh medium to each new culture. Incubate in a humidified 37°C, 5% CO<sub>2</sub> incubator.

*If using 75-cm<sup>2</sup> culture flasks, add 9 ml medium per flask.*

*Some labs now use incubators with 5% CO<sub>2</sub> and 4% O<sub>2</sub>. The low oxygen concentration is thought to simulate the in vivo environment of cells and to enhance cell growth.*



7. If necessary, feed subconfluent cultures after 3 or 4 days by removing old medium and adding fresh 37°C medium.
8. Passage secondary culture when it becomes confluent by repeating steps 1 to 7, and continue to passage as necessary.

## PASSAGING CELLS IN SUSPENSION CULTURE

Passaging of suspension cultures is somewhat less complicated than passaging of monolayer cultures. Because the cells are suspended in medium rather than attached to a surface, it is not necessary to disperse them enzymatically before passaging. However, before passaging, cells must be maintained in culture by feeding every 2 to 3 days until they reach confluency (i.e., until the cells clump together in the suspension and the medium appears turbid when the flask is swirled).

### Materials

See Basic Protocol

**NOTE:** All incubations are performed in a humidified 37°C, 5% CO<sub>2</sub> incubator unless otherwise specified. Some media (e.g., DMEM) may require altered levels of CO<sub>2</sub> to maintain pH 7.4.

**NOTE:** All solutions and equipment coming into contact with cells must be sterile, and proper aseptic technique should be used accordingly.

1. Feed cells as follows every 2 to 3 days until the cultures are confluent:
  - a. Remove flask of suspension cells from incubator, taking care not to disturb those that have settled to the flask bottom.
  - b. Aseptically remove and discard about one-third of the medium from the flask and replace with an equal volume of prewarmed (37°C) medium. If the cells are growing rapidly, add an additional 10% medium by volume in order to maintain an optimum concentration of  $1 \times 10^6$  cells/ml. Gently swirl flask to resuspend cells.

*Additional medium should be added to the cultures to provide enough nutrients to support the growth of the high number of cells.*

- c. Return flask to incubator. If there is <15 ml of medium in the flask, incubate flask in horizontal position to enhance cell/medium contact.

*At higher volumes of medium, the flask can be incubated in the vertical position.*

*If using a 25-cm<sup>2</sup> flask, there should be 20 to 30 ml of medium in the flask at confluency.*

2. On the days cultures are not being fed, check by swirling flask to resuspend cells and observing color changes.

*Acidic by-products are generated as a result of cell metabolism. The change in color of medium from pink to pink/yellow indicates a more acidic pH associated with active cell metabolism.*

3. When cultures are confluent ( $\sim 2.5 \times 10^6$  cells/ml), passage culture as follows:

- a. Remove flask from incubator and swirl flask so that cells are evenly distributed in the medium.
  - b. Aseptically remove half of the volume of cell suspension and place into a fresh flask, retaining the other half of the cell suspension in the original flask.

*Alternatively, the entire cell suspension can be removed from the original flask and divided equally into two fresh flasks. The original flask can be discarded, or if there is concern about the need for additional cells, the original flask can be retained and fed in an attempt to salvage any residual cells.*

## ALTERNATE PROTOCOL 1

### Commonly Used Techniques

#### A.3B.9

- c. Feed each flask with 7 to 10 ml prewarmed medium and return flasks to incubator.

*Some labs prefer to split the cells 1/3 or 1/4, although increasing the split ratio will result in a longer interval before subcultures reach confluency.*

## **FREEZING HUMAN CELLS GROWN IN MONOLAYER CULTURES**

It is sometimes desirable to store cell lines for future study. To preserve cells, avoid senescence, reduce the risk of contamination, and minimize effects of genetic drift, cell lines may be frozen for long-term storage. Without the use of a cryoprotective agent, freezing would be lethal to the cells in most cases. Generally, a cryoprotective agent such as dimethylsulfoxide (DMSO) is used in conjunction with complete medium for preserving cells at  $-70^{\circ}\text{C}$  or lower. DMSO acts to reduce the freezing point and allows a slower cooling rate. Gradual freezing reduces the risk of ice crystal formation and cell damage.

### **Materials**

Log-phase monolayer culture of cells in petri plate

Complete medium

Freezing medium: complete medium (e.g., complete DMEM or RPMI; see recipes) supplemented with 10% to 20% (v/v) FBS and 5% to 10% (v/v) DMSO,  $4^{\circ}\text{C}$

Benchtop clinical centrifuge with  $45^{\circ}$  fixed-angle or swinging-bucket rotor

**NOTE:** All incubations are performed in a humidified  $37^{\circ}\text{C}$ , 5%  $\text{CO}_2$  incubator unless otherwise specified. Some media (e.g., DMEM) may require altered levels of  $\text{CO}_2$  to maintain pH 7.4.

**NOTE:** All solutions and equipment coming into contact with cells must be sterile, and proper aseptic technique should be used accordingly.

1. Trypsinize cells from plate (see Basic Protocol, steps 1 to 4).

*It is best to use cells in log-phase growth for cryopreservation.*

2. Transfer cell suspension to a sterile centrifuge tube and add 2 ml complete medium with serum. Centrifuge 5 min at  $300$  to  $350 \times g$  ( $\sim 1500$  rpm in Fisher Centrifuge rotor), room temperature.

*The benchtop centrifuge can be accessorized depending on the anticipated volume. A variety of brands are available, including Eppendorf, Thermo Electron, and Beckman, and can be obtained from Fisher Scientific, VWR, and other laboratory equipment vendors.*

*Cells from three or more dishes from the same subculture can be combined in one tube.*

3. Remove supernatant and resuspend pellet in ml of  $4^{\circ}\text{C}$  freezing medium.
4. Add 4 ml of  $4^{\circ}\text{C}$  freezing medium, mix cells thoroughly, and place on wet ice.
5. Count cells using a hemacytometer (see Support Protocol 3). Dilute with more freezing medium as necessary to get a final cell concentration of  $10^6$  or  $10^7$  cells/ml.

*To freeze cells from a nearly confluent  $25\text{-cm}^2$  flask, resuspend in roughly 3 ml freezing medium.*

6. Pipet 1-ml aliquots of cell suspension into labeled 2-ml cryovials. Tighten caps on vials.
7. Place vials 1 hr to overnight in a  $-70^{\circ}\text{C}$  freezer, then transfer to a liquid nitrogen storage freezer.

*Keep accurate records of the identity and location of cells stored in liquid nitrogen freezers. Cells may be stored for many years and proper information is imperative for locating a particular line for future use.*

## **FREEZING CELLS GROWN IN SUSPENSION CULTURE**

Freezing cells from suspension culture is similar in principle to freezing cells from monolayer. The major difference is that suspension cultures need not be trypsinized.

### **Materials**

Cells in suspension culture  
Centrifuge tubes, sterile

**NOTE:** All incubations are performed in a humidified 37°C, 5% CO<sub>2</sub> incubator unless otherwise specified. Some media (e.g., DMEM) may require altered levels of CO<sub>2</sub> to maintain pH 7.4.

**NOTE:** All solutions and equipment coming into contact with cells must be sterile, and proper aseptic technique should be used accordingly.

1. Transfer cell suspension to a centrifuge tube and centrifuge 10 min at 300 to 350 × g (~1500 rpm in Fisher Centrifuge), room temperature.
2. Remove supernatant and resuspend pellet in 4°C freezing medium at a density of 10<sup>6</sup> to 10<sup>7</sup> cells/ml.

*Some laboratories freeze lymphoblastoid lines at the higher cell density because they plan to recover them in a larger volume of medium and because there may be a greater loss of cell viability upon recovery as compared to other types of cells (e.g., fibroblasts).*

3. Transfer 1-ml aliquots of cell suspension into cryovials and freeze as for monolayer cultures.

## **THAWING AND RECOVERING HUMAN CELLS**

When cryopreserved cells are needed for study, they should be thawed rapidly and plated at high density to optimize recovery.

### **Materials**

Cryopreserved cells stored in liquid nitrogen freezer  
70% (v/v) ethanol  
Complete medium (e.g., DMEM or RPMI; see recipes) containing 10% to 20% FBS (see recipe), 37°C

**CAUTION:** Protective clothing, particularly insulated gloves and goggles, should be worn when removing frozen vials or ampules from the liquid nitrogen freezer. The room containing the liquid nitrogen freezer should be well-ventilated. Care should be taken not to spill liquid nitrogen on the skin.

**NOTE:** All incubations are performed in a humidified 37°C, 5% CO<sub>2</sub> incubator unless otherwise specified. Some media (e.g., DMEM) may require altered levels of CO<sub>2</sub> to maintain pH 7.4.

**NOTE:** All solutions and equipment coming into contact with cells must be sterile, and proper aseptic technique should be used accordingly.

1. Remove vial from liquid nitrogen freezer and immediately place it into a 37°C water bath. Agitate vial continuously until medium is thawed.

*The medium usually thaws in <60 sec. Cells should be thawed as quickly as possible to prevent formation of ice crystals that can cause cell lysis. Try to avoid getting water around the cap of the vial.*

2. Wipe top of vial with 70% ethanol before opening.

*Some labs prefer to submerge the vial in 70% ethanol and air dry before opening.*

## **ALTERNATE PROTOCOL 2**

## **SUPPORT PROTOCOL 2**

### **Commonly Used Techniques**

## **A.3B.11**

3. Transfer thawed cell suspension into a sterile centrifuge tube containing 2 ml warm complete medium containing 20% FBS. Centrifuge 10 min at 150 to 200 × g (~1000 rpm in Fisher Centrifuge), room temperature. Discard supernatant.

*Cells are washed with fresh medium to remove residual DMSO.*

4. Gently resuspend cell pellet in small amount (~1 ml) of complete medium/20% FBS and transfer to properly labeled culture plate containing the appropriate amount of medium.

*Cultures are reestablished at a higher cell density than that used for original cultures because there is some cell death associated with freezing. Generally, 1 ml cell suspension is reseeded in 5 to 20 ml medium.*

5. Check cultures after ~24 hr to ensure that cells have attached to the plate.
6. Change medium after 5 to 7 days or when pH indicator (e.g., phenol red) in medium changes color. Keep cultures in medium with 20% FBS until cell line is reestablished.

*If recovery rate is extremely low, only a subpopulation of the original culture may be growing; be extra careful of this when working with cell lines known to be mosaic.*

### **SUPPORT PROTOCOL 3**

## **DETERMINING CELL NUMBER AND VIABILITY WITH A HEMACYTOMETER AND TRYPAN BLUE STAINING**

Determining the number of cells in culture is important in standardization of culture conditions and in performing accurate quantitation experiments. A hemacytometer is a thick glass slide with a central area designed as a counting chamber. Cell suspension is applied to a defined area and counted so cell density can be calculated.

The exact design of the hemacytometer may vary; the one described here is the Improved Neubauer from VWR (Fig. A.3B.1). The central portion of the slide is the counting platform, which is bordered by a 1-mm groove. The central platform is divided into two counting chambers by a transverse groove. Each counting chamber consists of a silver footplate on which is etched a 3 × 3-mm grid. This grid is divided into nine secondary squares, each 1 × 1 mm. The four corner squares and the central square are used for determining the cell count. The corner squares are further divided into 16 tertiary squares and the central square into 25 tertiary squares to aid in cell counting.


Accompanying the hemacytometer slide is a thick, even-surfaced coverslip. Ordinary coverslips may have uneven surfaces, which can introduce errors in cell counting; therefore, it is imperative that the coverslip provided with the hemacytometer be used in determining cell number.

### **Materials**

70% (v/v) ethanol  
Cell suspension  
0.4% (w/v) trypan blue or 0.4% (w/v) nigrosin, prepared in HBSS (See recipe)  
Hemacytometer with coverslip (Improved Neubauer, Baxter Scientific)  
Hand-held counter

### **Prepare hemacytometer**

1. Clean surface of hemacytometer slide and coverslip with 70% alcohol.  
*Coverslip and slide should be clean, dry, and free from lint, fingerprints, and watermarks.*
2. Wet edge of coverslip slightly with tap water and press over grooves on hemacytometer. The coverslip should rest evenly over the silver counting area.



**Figure A.3B.1** Hemacytometer slide (Improved Neubauer) and coverslip. Coverslip is applied to slide and cell suspension is added to counting chamber using a Pasteur pipet. Each counting chamber has a  $3 \times 3$ -mm grid (enlarged). The four corner squares (1, 2, 4, and 5) and the central square (3) are counted on each side of the hemacytometer (numbers added).

#### ***Prepare cell suspension***

3. For cells grown in monolayer cultures, detach cells from surface of dish using trypsin (see Basic Protocol).
4. Dilute cells with tissue culture medium as needed to obtain a uniform suspension. Disperse any clumps.

*When using the hemacytometer, a maximum cell count of 20 to 50 cells per 1-mm square is recommended.*

#### ***Load hemacytometer***

5. Use a sterile Pasteur pipet to transfer cell suspension to edge of hemacytometer counting chamber. Hold tip of pipet under the coverslip and dispense one drop of suspension.

*Suspension will be drawn under the coverslip by capillary action.*

*Hemocytometer should be considered nonsterile. If cell suspension is to be used for cultures, do not reuse the pipet and do not return any excess cell suspension in the pipet to the original suspension.*

6. Fill second counting chamber.

#### **Count cells**

7. Allow cells to settle for a few minutes before beginning to count. Blot off excess liquid.
8. View slide on microscope with 100× magnification.

*A 10× ocular with a 10× objective = 100× magnification.*

*Position slide to view the large central area of the grid (section 3 in Fig. A.3B.1); this area is bordered by a set of three parallel lines. The central area of the grid should almost fill the microscope field. Subdivisions within the large central area are also bordered by three parallel lines and each subdivision is divided into sixteen smaller squares by single lines. Cells within this area should be evenly distributed without clumping. If cells are not evenly distributed, wash and reload hemacytometer.*

9. Use a hand-held counter to count cells in each of the four corner and central squares (Fig. A.3B.1, squares numbered 1 to 5). Repeat counts for other counting chamber.

*Five squares (four corner and one center) are counted from each of the two counting chambers for a total of ten squares counted.*

*Count cells touching the middle line of the triple line on the top and left of the squares. Do not count cells touching the middle line of the triple lines on the bottom or right side of the square.*

#### **Calculate cell number**

10. Determine cells per ml by the following calculations:

cells/ml = average count per square × dilution factor ×  $10^4$   
total cells = cells/ml × total original volume of cell suspension from which sample was taken.

*The volume correction factor for the hemacytometer is  $10^4$ : each square is  $1 \times 1$  mm and the depth is 0.1 mm.*

#### **Stain cells with trypan blue to determine cell viability**

11. Determine number of viable cells by adding 0.5 ml of 0.4% trypan blue, 0.3 ml HBSS, and 0.1 ml cell suspension to a small tube. Mix thoroughly and let stand 5 min before loading hemacytometer.

*Either 0.4% trypan blue or 0.4% nigrosin can be used to determine the viable cell number. Nonviable cells will take up the dye, whereas live cells will be impermeable to it.*

12. Count total number of cells and total number of viable (unstained) cells. Calculate percent viable cells as follows:

$$\% \text{ viable cells} = \frac{\text{number of unstained cells}}{\text{total number of cells}} \times 100$$

13. Decontaminate coverslip and hemacytometer by rinsing with 70% ethanol and then deionized water. Air dry and store for future use.

## PREPARING CELLS FOR TRANSPORT

Both monolayer and suspension cultures can easily be shipped in 25-cm<sup>2</sup> tissue culture flasks. Cells are grown to near confluency in a monolayer or to desired density in suspension. Medium is removed from monolayer cultures and the flask is filled with fresh medium. Fresh medium is added to suspension cultures to fill the flask. *It is essential that the flasks be completely filled with medium to protect cells from drying if flasks are inverted during transport.* It is also imperative that the flasks have nonvented caps. The cap is tightened and taped securely in place. The flask is sealed in a leak-proof plastic bag or other leak-proof container designed to prevent spillage in the event that the flask should become damaged. The primary container is then placed in a secondary insulated container to protect it from extreme temperatures during transport. A biohazard label is affixed to the outside of the package. Generally, cultures are transported by same-day or overnight courier.

Cells can also be shipped frozen. The vial containing frozen cells is removed from the liquid nitrogen freezer and placed immediately on dry ice in an insulated container to prevent thawing during transport.

## REAGENTS AND SOLUTIONS

*Use deionized, distilled water in all recipes and protocol steps. Use double- or triple-distilled water or commercially available tissue culture water for media preparation. Suitable suppliers for media and components include Invitrogen, Cambrex, HyClone, and Sigma-Aldrich. For common stock solutions, see APPENDIX 2A; for suppliers, see SUPPLIERS APPENDIX.*

### Complete DMEM

*Dulbecco's modified Eagle medium, high-glucose formulation (e.g., Invitrogen), containing:*

5%, 10%, or 20% (v/v) FBS (optional; see recipe)

1% (v/v) nonessential amino acids

2 mM L-glutamine (see recipe)

100 U/ml penicillin

100 µg/ml streptomycin sulfate

Filter sterilize and store ≤1 month at 4°C

*The percentage of serum (usually fetal bovine serum) used in a protocol step is indicated by a numeral hyphenated to the base medium name. Thus, "complete DMEM-10" indicates that 10% FBS is used. Absence of a numeral indicates that no serum is used.*

*DMEM containing 4500 mg/liter D-glucose can be obtained from Invitrogen. DMEM is also known as Dulbecco's minimum essential medium.*

### Complete RPMI

*RPMI 1640 medium (e.g., Invitrogen) containing:*

2%, 5%, 10%, 15%, or 20% FBS (optional; see recipe)

2 mM L-glutamine (see recipe)

100 U/ml penicillin

100 µg/ml streptomycin sulfate

Filter sterilize and store ≤1 month at 4°C

### **FBS (fetal bovine serum)**

Thaw purchased fetal bovine serum (shipped on dry ice and kept frozen at  $-20^{\circ}\text{C}$  until needed). Store 3 to 4 weeks at  $4^{\circ}\text{C}$ . If FBS is not to be used within this time, aseptically divide into smaller aliquots and refreeze until used. Store  $\leq 1$  year at  $-20^{\circ}\text{C}$ .

*Repeated thawing and refreezing should be avoided as it may cause denaturation of the serum.*

*In some cases, heat inactivation may be warranted (see Culture Medium Preparation). To inactivate FBS, heat 30 to 60 min in a  $56^{\circ}\text{C}$  water bath. Alternatively, FBS may be inactivated through radiation treatment.*

### **L-Glutamine, 0.2 M (100 $\times$ )**

Thaw frozen L-glutamine, aliquot aseptically into usable portions, then refreeze. For convenience, L-glutamine can be stored in 1-ml aliquots if 100-ml bottles of medium are used, and in 5-ml aliquots if 500-ml bottles are used. Store  $\leq 1$  year at  $-20^{\circ}\text{C}$ .

*Many laboratories supplement medium with 2 mM L-glutamine—1% (v/v) of 100 $\times$  stock—just prior to use.*

### **HBSS (Hank's balanced salt solution)**

5.4 mM KCl  
0.3 mM  $\text{Na}_2\text{HPO}_4$   
0.4 mM  $\text{KH}_2\text{PO}_4$   
4.2 mM  $\text{NaHCO}_3$   
1.3 mM  $\text{CaCl}_2$   
0.5 mM  $\text{MgCl}_2$   
0.6 mM  $\text{MgSO}_4$   
137 mM NaCl

5.6 mM D-glucose

0.02% phenol red (optional)

Add  $\text{H}_2\text{O}$  to 1 liter and adjust pH to 7.4 with 1 M HCl or 1 M NaOH

*HBSS can be purchased from Biofluids or BioWhittaker.*

*HBSS may be made or purchased without  $\text{CaCl}_2$  and  $\text{MgCl}_2$ . These are optional components that usually have no effect on an experiment. In some cases, however, their presence may be detrimental to a procedure. Consult the individual protocol to see if the presence or absence of these components is recommended in the materials list.*

### **Trypsin/EDTA solution**

*Prepare in sterile HBSS (See recipe) or 0.9% (w/v) NaCl:*

0.25% (w/v) trypsin

0.2% (w/v) EDTA

Store  $\leq 1$  year (until needed) at  $-20^{\circ}\text{C}$

*Most laboratories prefer to purchase trypsin/EDTA as a prepared solution, which is available from vendors including Sigma-Aldrich, Invitrogen, and Cambrex. This is a convenient and cost-effective alternative to preparing the solution within the laboratory. Trypsin/EDTA solution is available in various concentrations including 10 $\times$ , 1 $\times$ , and 0.25% (w/v). It is received frozen from the manufacturer and can be thawed and aseptically aliquotted into smaller volumes. Specific applications may require different concentrations of trypsin; the appropriate methods should be consulted for details.*

*EDTA (disodium ethylenediamine tetraacetic acid) is added as a chelating agent to bind  $\text{Ca}^{2+}$  and  $\text{Mg}^{2+}$  ions that can interfere with the action of trypsin.*



## COMMENTARY

### Background Information

At its inception in the early twentieth century, tissue culture was applied to the study of tissue fragments in culture. New growth in culture was limited to cells that migrated out from the initial tissue fragment. Tissue culture techniques evolved rapidly, and since the 1950s culture methods have allowed the growth and study of dispersed cells in culture (Freshney, 1996). Cells dispersed from the original tissue can be grown in monolayers and passaged repeatedly to give rise to a relatively stable cell line.

Four distinct growth stages have been described for primary cells maintained in culture. First, cells adapt to the *in vitro* environment. Second, cells undergo an exponential growth phase lasting through ~30 passages. Third, the growth rate of cells slows, leading to a progressively longer generation time. Finally, after 40 or 50 passages, cells begin to senesce and die (Lee, 1991). It may be desirable to study a particular cell line over several months or years, so monolayer cultures can be preserved to retain the integrity of the cell line. Aliquots of early-passage cell suspensions are frozen, then thawed, and cultures reestablished as needed. Freezing monolayer cultures prevents changes due to genetic drift and avoids loss of cultures due to senescence or accidental contamination (Freshney, 1996).

Cell lines are commercially available from a number of sources, including the American Type Culture Collection (ATCC; <http://www.atcc.org>) and the Human Genetic Mutant Cell Repository at the Coriell Cell Repository (CCR; <http://www.coriell.org>). These cell repositories are a valuable resource for researchers who do not have access to suitable patient populations. An up-to-date listing of available cell lines will be provided upon request. The cell lines are preserved in liquid nitrogen, and information on the characteristics are supplied by the distributor. Immortalized cell lines can be maintained in culture indefinitely, while nontransformed cells have a limited life-span *in vitro*. For human fibroblasts, about 20 population doublings occur between the initial growth in the primary culture and the first subculture (Priest, 1997). About 50 to 60 population doublings is the standard rule for maintaining fibroblast cell lines in culture before risking the loss of cell lines through senescence, cytogenetic or biochemical changes, or contamination. If it is anticipated that these cells will be needed for future

study, cells should be frozen at an early culture stage and retrieved for future use (Verma and Babu, 1995).

### Critical Parameters

Use of aseptic technique is essential for successful tissue culture. Cell cultures can be contaminated at any time during handling, so precautions must be taken to minimize the chance of contamination. All supplies and reagents that come into contact with cultures must be sterile, and all work surfaces should be kept clean and free from clutter.

Cultures should be 75% to 100% confluent when selected for subculture. Growth in monolayer cultures will be adversely affected if cells are allowed to become overgrown. Passaging cells too early will result in a longer lag time before subcultures are established. Following dissociation of the monolayer by trypsinization, serum or medium containing serum should be added to the cell suspension to stop further action by trypsin that might be harmful to cells.

When subculturing cells, add a sufficient number of cells to give a final concentration of  $\sim 5 \times 10^4$  cells/ml in each new culture of human cell lines. Optimal concentrations are cell-type specific. Package inserts provided with commercially available cell lines contain information on recommended cell density, culture media, and appropriate culture temperatures. Cells plated at too low a density may be inhibited or delayed in entry into growth stage. Cells plated at too high a density may reach confluence before the next scheduled subculturing; this could lead to cell loss and/or cessation of proliferation. The growth characteristics for different cell lines vary. A lower cell concentration ( $10^4$  cells/ml) may be used to initiate subcultures of rapidly growing cells, and a higher cell concentration ( $10^5$  cells/ml) may be used to initiate subcultures of more slowly growing cells. Adjusting the initial cell concentration permits establishment of a regular, convenient schedule for subculturing—e.g., once or twice a week (Freshney, 1996).

Cells in culture will undergo changes in growth, morphology, and genetic characteristics over time. Such changes can adversely affect reproducibility of laboratory results. Non-transformed cells will undergo senescence and eventual death if passaged indefinitely. The time of senescence will vary with cell line, but generally at between 40 and 50 population doublings fibroblast cell lines begin to

senesce. Cryopreservation of cell lines will protect against these adverse changes and will avoid potential contamination.

Cultures selected for cryopreservation should be in log-phase growth and free from contamination. Cells should be frozen at a concentration of  $10^6$  to  $10^7$  cells/ml. Cells should be frozen gradually and thawed rapidly to prevent formation of ice crystals that may cause cells to lyse. Cell lines can be thawed and recovered after long-term storage in liquid nitrogen. The top of the freezing vial should be cleaned with 70% alcohol before opening to prevent introduction of contaminants. To aid in recovery of cultures, thawed cells should be reseeded at a higher concentration than that used for initiating primary cultures. Careful records regarding identity and characteristics of frozen cells as well as their location in the freezer should be maintained to allow for easy retrieval.

For accurate cell counting, the hemacytometer slide should be clean, dry, and free from lint, scratches, fingerprints, and watermarks. The coverslip supplied with the hemacytometer should always be used because it has an even surface and is specially designed for use with the counting chamber. Use of an ordinary coverslip may introduce errors in cell counting. If the cell suspension is too dense or the cells are clumped, inaccurate counts will be obtained. If the cell suspension is not evenly distributed over the counting chamber, the hemacytometer should be washed and reloaded.

### Anticipated Results

Confluent cell lines can be successfully subcultured in the vast majority of cases. The yield of cells derived from monolayer culture is directly dependent on the available surface area of the culture vessel (Freshney, 1996). Overly confluent cultures or senescent cells may be difficult to trypsinize, but increasing the time of trypsin exposure will help dissociate resistant cells. Cell lines can be propagated to get sufficient cell populations for cytogenetic, biochemical, and molecular analyses.

It is well accepted that anyone can successfully freeze cultured cells; it is thawing and recovering the cultures that presents the problem. Cultures that are healthy and free from contamination can be frozen and stored indefinitely. Cells stored in liquid nitrogen can be successfully thawed and recovered in over 95% of cases. Several aliquots of each cell line should be stored to increase the chance of recovery. Cells should be frozen gradually,

with a temperature drop of  $\sim 1^\circ\text{C}$  per minute, but thawed rapidly. Gradual freezing and rapid thawing prevents formation of ice crystals that might cause cell lysis.

Accurate cell counts can be obtained using the hemacytometer if cells are evenly dispersed in suspension and free from clumps. Determining the proportion of viable cells in a population will aid in standardization of experimental conditions.

### Time Considerations

Establishment and maintenance of mammalian cell cultures require a regular routine for preparation of media and feeding and passaging cells. Cultures should be inspected regularly for signs of contamination and to determine if the culture needs feeding or passaging.

### Literature Cited

- Barch, M.J., Lawce, H.J., and Arsham, M.S. 1991. Peripheral Blood Culture. *In* The ACT Cytogenetic Laboratory Manual, 2nd ed. (M.J. Barch, ed.) pp. 17-30. Raven Press, New York.
- Freshney, R.I. 1996. Culture of Animal Cells. A Manual of Basic Techniques, 3rd ed. Wiley-Liss, New York.
- Hyclone Labs. 1996. Heat inactivation: Are you wasting your time? *Art to Science* 15:1-5.
- Knutsen, T. 1991. The ACT Cytogenetic Laboratory Manual, 2nd ed. (M.J. Barch, ed.) pp. 563-587. Raven Press, New York.
- Lee, E.C. 1991. Cytogenetic Analysis of Continuous Cell Lines. *In* The ACT Cytogenetic Laboratory Manual, 2nd ed. (M.J. Barch, ed.) pp. 107-148. Raven Press, New York.
- Newman, C. 2003. Serum-free cell culture—the ethical, scientific and economic choice. *The Biomedical Scientist*, pp. 941-942.
- Priest, J.H. 1997. General cell culture principles and fibroblast culture. *In* The AGT Cytogenetics Laboratory Manual, 3rd ed. (M.J. Barch, T. Knutsen, and J.L. Spurbeck, eds.). Lippincott-Raven, Philadelphia.
- Rooney, D.E. and Czepulskiowski, B.H. (eds.). 2001. Human Cytogenetics: Constitutional Analysis: A Practical Approach, 3rd ed. Oxford University Press, New York.
- Triglia, R.P. and Linscott, W.D. 1980. Titers of nine complement components, conglutinin and C3b inactivator in adult and fetal bovine sera. *Mol. Immunol.* 17:741-748.
- Verma, R.S. and Babu, A. 1995. Human Chromosomes: Principles and Techniques, 2nd ed. McGraw-Hill, New York.

### Key Reference

- Lee, 1991. See above.  
*Contains pertinent information on cell culture requirements, including medium preparation and sterility. Also discusses trypsinization, freezing and thawing, and cell counting.*

## Internet Resources

<http://www.unc.edu/depts/tcf/info.html>

*This site contains troubleshooting information for treating tissue culture contamination, including discussion of contamination risks and recommendations.*

<http://www.cdc.gov/od/ohs/pdffiles/BSC-3.pdf>

*Primary Containment for Biohazards: Selection, Installation, and Use of Biological Safety Cabinets, 2nd ed. Sept 2001. U.S. Dept of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention and National Institutes of Health, U.S. Government Printing Office, Washington.*

<http://www.cdc.gov/od/ohs/symp5/jyrtext.htm>

*The 1, 2, 3's of Biosafety Levels by J.Y. Richmond. Adapted from Richmond, J.Y. and McKinney, R.W. 1993. Biosafety in Microbiological and Biomedical Laboratories, 3rd ed. U.S. Department of Health*

*and Human Services, CDC/NIH, U.S. Government Printing Office, Washington, D.C.*

<http://www.ehs.cornell.edu/bio/cabinets.htm>

*This site contains a comprehensive discussion of the use of biological safety cabinets, including operational procedures and the advantages and disadvantages of ultraviolet lights.*

<http://www.cdc.gov/od/ohs/biosfty/bmbl4/b4aa.htm>

*This site contains information pertaining to design, function, and use of biological safety cabinets, including comparison of Class I, II, and III safety cabinets.*

---

Contributed by Mary C. Phelan  
Molecular Pathology Laboratory Network  
Maryville, Tennessee